

## WARRANTY

The warranty for the Pediatric SACH Foot is one year. The use of WillowWood pediatric components with any manufacturer's adult-sized components will void the warranty. Use of the Pediatric SACH Foot for amputees whose modified body weight is more than 132 lbs (60 kg) or who engage in extremely high and abusive activity is against WillowWood's recommendations and will void the one year warranty. Modified body weight is defined as the weight of the amputee plus any loads carried by the amputee. "Extremely high and abusive activities" are defined as activities such as skydiving, karate, and judo; activities that could result in injury to an individual's natural feet; and activities that expose the prosthesis to corrosives such as salt water.

## WARRANTY DISCLAIMER

WillowWood warrants that each product manufactured will, at the time of delivery, be of workmanlike quality and substantially free of defects. WILLOWWOOD MAKES NO OTHER WARRANTY, IMPLIED, OR EXPRESSED, AND MAKES NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. This warranty shall terminate immediately upon an action to combine our products with other materials or in any manner to change the nature of our products. The sole remedy is replacement of the products or credit for the products. WillowWood's liability shall not exceed the purchase price of the product. WillowWood shall not be liable for any indirect, incidental, or consequential damage.

## WILLOWWOOD RETENTION OF RIGHTS

WillowWood retains all intellectual property rights reflected or incorporated in its physical products, regardless of the transfer of the physical products to another party or parties.

## PARTS AND ACCESSORIES

### Endoskeletal Accessories

PPF-1208 Pediatric Titanium Foot Pyramid Adapter  
PPF-1209 Pediatric Aluminum Foot Pyramid Receiver

### Exoskeletal Accessories

PPF-1540 Pediatric Exoskeletal Laminating Core  
PPI-1518SL Pediatric Impulse Exoskeletal Ankle Block (Size 15-18) Small Left  
PPI-1518SR Pediatric Impulse Exoskeletal Ankle Block (Size 15-18) Small Right  
PPI-1922LL Pediatric Impulse Exoskeletal Ankle Block (Size 19-22) Large Left  
PPI-1922LR Pediatric Impulse Exoskeletal Ankle Block (Size 19-22) Large Right

**Pediatric Growth Kit** PPF-GKIT

**Pediatric Titanium Symes Nut** PPF-FS006

# WillowWood®

## Pediatric SACH Foot Instructions



### WHAT'S IN THE BOX

Pediatric SACH Foot  
M8 x 32 Foot Bolt (700-B202)  
M8 x 24 Foot Bolt (700-B200)  
Set of 3 Heel Inserts (foot sizes 15-22 only)  
Instructions  
Patient Advisory

# WillowWood®

The Ohio Willow Wood Company  
15441 Scioto Darby Road  
Mt. Sterling, Ohio 43143 USA  
Phone: 740.869.3377 or 800.848.4930  
Fax: 740.869.4374 www.willowwoodco.com



Ohio Willow Wood Company B.V.  
Keizersgracht 62/64  
1015 CS Amsterdam  
The Netherlands



**Notice: WillowWood pediatric components have been tested and approved for use only with other WillowWood pediatric components. The use of WillowWood pediatric components with any manufacturer's adult-sized components will void the warranty. The use of WillowWood pediatric components with another manufacturer's pediatric components is done so at the sole judgment of the attending prosthetist.**

## FOOT BOLTS

WillowWood recommends using the M8 x 32 mm bolt (Part No. 700-B202) when connecting to a Pediatric Titanium Foot Pyramid Adapter, or the M8 x 24 mm bolt (Part No. 700-B200) when connecting to a Pediatric Titanium Foot Pyramid Receiver, Symes Nut, or Exoskeletal Laminating Core. The use of fasteners other than those provided for this product is done so at the sole judgement of the attending prosthetist. The use of the provided fastener is not recommended for any product other than WillowWood feet.

To significantly reduce the possibility of bolt failure, the structures connected by the bolt must both be rigid. (Non-rigid structures demonstrate slight deformation upon loading, resulting in bending of the foot bolt that inevitably leads to failure.) The Pediatric SACH Foot has been thoroughly tested to ensure that no keel deformation will occur; the prosthetist is then responsible for selecting a component to attach to the keel that is at least as rigid as the keel itself. The foot bolt is a high quality capscrew that has been manufactured especially for this purpose, but WillowWood cannot guarantee its suitability for use with components from other manufacturers.

## INSTALLATION

**Foot Pyramid:** Use the Pediatric Titanium Foot Pyramid Adapter (Part No. PPF-1208) or Aluminum Foot Pyramid Receiver (Part No. PPF-1209) for best results. The proximal surface of the keel has been specially designed to accommodate these adapters. When assembled, the end of the foot bolt must be within 2 mm of the top of the foot pyramid. (Please note that if the foot bolt extends more than 2 mm above the foot pyramid, the bolt will interfere with the function of WillowWood's Pediatric components.)

The proximal surface also accommodates the spacers in the Pediatric Growth Kit (Part No. PPF-GKIT), which may be used to increase the length of the prosthesis up to 7/8".

If using another manufacturer's foot pyramid, an interface plate may be required to accommodate the keyways on the bottom of the foot pyramid.

**Torque:** Tighten the Foot Bolt to 15 ft-lbs (20 Nm).

**Heel Inserts:** for foot sizes 15-22, install heel inserts by pressing them into the recess on the bottom surface of the foot. No adhesive is necessary for definitive use. Do not use the foot without a heel insert in place.

Sizes 9-14 do not include heel inserts.

**Exoskeletal Applications:** To fabricate an exoskeletal prosthesis with a Pediatric SACH foot, use the Pediatric Exoskeletal Laminating Core (Part No. PPF-1540) or the Pediatric SACH Exoskeletal Ankle Block and follow the directions included in the kit.

## MODIFYING THE FOOT

The urethane foam shell may be lightly sanded with little resulting loss of strength. Do not expose the internal components of the foot, as this could reduce the structural integrity of the foot.

## EXPOSURE TO WATER

The Pediatric SACH Foot is essentially water resistant, provided no open cracks have developed and no area of the foot has been sanded. Seal the bolt hole with silicone before exposing the foot to water. Complete submersion in water is not recommended.

## PATIENT ADVISORY WARNING

The attached Patient Advisory Warning enables you, the prosthetist, to effectively notify your patients of the limitations of the components in their prosthesis, and of the need to monitor their weight and activity levels. Please review the Patient Advisory Warning with the patient upon delivery of a prosthesis with a Pediatric SACH Foot. The patient and the prosthetist should then sign the Patient Advisory Warning to acknowledge that it has been reviewed and understood by both parties. Give one signed copy to the patient and place one copy in the patient's file.

If a patient's weight or activity level increases after receiving a prosthesis with a Pediatric SACH Foot, the patient should immediately contact the prosthetist to determine whether replacement components are necessary. If a patient continues to use a prosthesis with a Pediatric SACH Foot after experiencing an increase in weight and/or activity level, the foot could fail with the possibility of serious injury to the patient.

To ensure that the correct components are selected for each patient, the prosthetist should weigh the patient on scales in the prosthetist's office. Do not rely on the patient's estimate of his/her own weight. Instruct the patient to monitor his/her weight weekly to ensure that it remains in a range appropriate for the prosthetic components being used.